

Aquasil Easy Mix Putty Recall (7/16)

Reason/Information:

DENTSPLY Caulk, Inc. Issued a Voluntary Market Recall Notice on the following materiel (See Image on Message under Additional Documentation/Attachment). **Reason:** DENTSPLY received reports where two lots of Aquasil Easy Mix Putty were not setting.

Disposition/Instructions:

Actions to be taken:

--Examine current inventory.

--Provide DENTSPLY with the requested information by faxing the Consignee Certification form (See Image on Message under Additional Documentation/Attachment) to 302-422-7306 (Attention Customer Quality Department). DENTSPLY will need confirmation that customers have inventory from the recalled lot or that customers do not have any of the recalled lot.

--As soon as DENTSPLY receives the requested information and returned materiel, if applicable, the Customer Quality Department will coordinate for replacement materiel.

If customers have questions regarding this notification contact DENTSPLY Caulk, Inc. at 800-532-2855 Ext. 57320, 57310 or 302-422-4511.

Item Information:

NSN (FSC-NIIN): 6520 - 016062790

PartNo: 678120

Nomenclature: KIT IMPRESSION MATERIAL HYROPHILIC VINYL POLY PUTTY GUN SCOOP CATALYST 320ML

UI: EA

Manufacturer: DENTSPLY INTERNATIONAL, INC.

Distributor:

LOT / SN:

150622 150629

Additional information and documentation can be found by clicking [here](#).