

0.9% Sodium Chloride Irrigation / Baxter Healthcare Corporation / Owens & Minor / Cardinal Health –RECALL (6/14)

Reason/Information:

Baxter issued an Urgent Product Recall on the following materiel (See Image on Message under Additional Documentation/Attachment). Reason: Due to particulate matter found inside the container.

Disposition/Instructions:

Baxter is requesting that customers take the following actions:

Locate and remove all affected product from customer facilities (the product code and lot number can be found on the individual product package or shipping carton).

Contact Baxter Healthcare Center for Service to arrange for return and credit. The Center for Service can be reached at 888-229-0001 between the hours of 7:00 am and 6:00 pm CDT. Have the customer's eight-digit Baxter ship-to account number ready when calling.

Complete a Customer Reply Form, and return it to Baxter by either fax or scanned e-mail.

For general questions or to obtain a Customer Reply Form, call 847-948-2000.

Note: Product shipped between 24 September 2013 and 14 February 2014.

Item Information:

NSN (FSC-NIIN): 6505-011137515
Nomenclature: 0.9% SODIUM CHLORIDE IRRIGATION; USP; PLASTIC POUR BOTTLE
CONTAINER; 1000ML
UI: EA
Manufacturer: BAXTER HEALTHCARE

LOT / SN:
G109272

NOTES:

NOTE 1:

"This product is NOT a direct match with the above listed NSN, however, the Medical Master Catalog indicates that it may be associated."

NSN (FSC-NIIN): 6505-011533787
Nomenclature: 0.9% SODIUM CHLORIDE IRRIGATION; USP; PLASTIC POUR BOTTLE
CONTAINER; 1000ML
UI: EA
Manufacturer: BAXTER HEALTHCARE

LOT / SN:

G109272

NOTES:

NOTE 1:

"This product is NOT a direct match with the above listed NSN, however, the Medical Master Catalog indicates that it may be associated."

NSN (FSC-NIIN): 6505-014624194

NDC: 00338-0048-04

Part Number: 2F7124

Nomenclature: 0.9% SODIUM CHLORIDE IRRIGATION; USP; PLASTIC POUR BOTTLE CONTAINER; 1000ML

UI: EA

Manufacturer: BAXTER HEALTHCARE

LOT / SN:

G109272

NOTES:

NOTE 1:

O&M SKU #07062F7124

NOTE 2:

AmerisorceBergen #10038549596692

NSN (FSC-NIIN): 6505-014716243

Nomenclature: 0.9% SODIUM CHLORIDE IRRIGATION; USP; PLASTIC POUR BOTTLE CONTAINER; 1000ML

UI: EA

Manufacturer: BAXTER HEALTHCARE

LOT / SN:

G109272

NOTES:

NOTE 1:

"This product is NOT a direct match with the above listed NSN, however, the Medical Master Catalog indicates that it may be associated."

POC Contact Information:

USAMMC-E

Email: usarmy.rheinland-pfalz.medcom-usammce.list.dtd-quality-assurance@mail.mil

Phone (COMM): 011-49-6331-86-7118/7181

Phone (DSN): 495-7118

Fax: 6218

US ARMY

Email: usarmy.detrick.medcom-usamma.mbx.doc@mail.mil

Phone (COMM): 301-619-4300
Phone (DSN): 343-4300
Fax: 301-619-4468

US AIR FORCE

Email: afmoa.sgalc@us.af.mil

Phone (COMM): 301-619-4170

Phone (DSN): 343-4170

Fax: 301-619-2557

US NAVY

Email: MMQC2@MED.NAVY.MIL

Phone (COMM): 301-619-3085

Phone (DSN): 343-3085

Fax: 301-619-2473

Additional Message Recipients:

Please ensure widest dissemination of information to all interested entities.

Message Dissemination Authorization:

US AIR FORCE

AF Activities Will Take Action As Prescribed In AFI 41-209, Medical Logistics Support, Chapters 3 And 9. For MAJCOMS & NGB--This Msg Has Been Transmitted To All Designated Subordinate Medical Activities.

US ARMY

See Army Regulation (Ar) 40-61, 28 January 2005, Chapter 4, And The Department Of The Army Supply Bulletin (Sb 8-75-11) For Applicable Policies And Procedures.

Additional Documentation may be found [here](#).