

Osseotite 2 Certain Prevail / Osseotite Parallel Walled Certain / Biomet 3i Recall (9/13)

Reason/Information:

Biomet 3i issued a recall on the following materiel, Osseotite 2 Certain Prevail 5mm doesn't have the internal hex feature. Osseotite Parallel Walled Certain 4mm doesn't have the internal thread feature.

Disposition/Instructions:

Customers were/will be contacted in writing via email, facsimile, postal mail or by telephone. Customer will be instructed to check their respective inventory for the affected materiel and provide feedback to Biomet 3i. Customers are to return the affected materiel to Biomet 3i. Biomet 3i will issue replacement materiel. General questions, call 561-776-6700.

Note: Materiel distributed from 06 October 2011 to 18 March 2013 and manufactured from 20 January 2011 to 30 September 2011.

Item Information:

NSN (FSC-NIIN): 6520-NS1
Part Number: XIIOS5411
Nomenclature: OSSEOTITE 2 CERTAIN PREVAIL 5MM(D)X 4.1MM (P) X X 11.5MM (L)
DENTAL IMPLANT
UI: EA
Manufacturer: BIOMET 3I

LOT / SN:
090 2011 247

NOTES:

NSN (FSC-NIIN): 6520-NS2
Part Number: IOSS415
Nomenclature: OSSEOTITE PARALLEL WALLED CERTAIN 4MM(D)X 15MM (L)
DENTAL IMPLANT
UI: EA
Manufacturer: BIOMET 3I

LOT / SN:
121 2010 294

NOTES:

Service/Additional Instructions:

CONSIGNEES:Lackland Wilford Medical Center

POC Contact Information:

USAMMC-E

Email: USAMMCE_QA@AMEDD.ARMY.MIL

Phone (COMM): 011-49-6331-86-7118/7181

Phone (DSN): 495-7118
Fax: 6218

US ARMY

Email: usarmy.detrick.medcom-usamma.mbx.doc@mail.mil
Phone (COMM): 301-619-4300
Phone (DSN): 343-4300
Fax: 301-619-4468

US AIR FORCE

Email: AFMOA.SGALC@DETRICK.AF.MIL
Phone (COMM): 301-619-4170
Phone (DSN): 343-4170
Fax: 301-619-2557

US NAVY

Email: MMQC2@MED.NAVY.MIL
Phone (COMM): 301-619-3085
Phone (DSN): 343-3085
Fax: 301-619-2473

Additional Message Recipients:

Please ensure widest dissemination of information to all interested entities.

Message Dissemination Authorization:

US AIR FORCE

AF Activities Will Take Action As Prescribed In AFI 41-209, Medical Logistics Support, Chapters 3 And 9. For MAJCOMS & NGB--This Msg Has Been Transmitted To All Designated Subordinate Medical Activities.

US ARMY

See Army Regulation (Ar) 40-61, 28 January 2005, Chapter 4, And The Department Of The Army Supply Bulletin (Sb 8-75-11) For Applicable Policies And Procedures.