

**Kodak INSIGHT and Ultra-Speed Dental Film / Carestream Health, Inc. RECALL (9/14)**

**Reason/Information:**

Carestream Health issued a recall on the following material. Reason: Carestream manufacturing identified an incorrect orientation of the indicator dot on dental intraoral film packet labeling. The Indicator Dot on the film packet's outer printed paper labeling was incorrectly positioned relative to the location of the raised dot on the film. This issue may cause some confusion for the dental practitioner if the processed film is placed on the view box by reference.

**Disposition/Instructions:**

Recall notices were mailed to customers via overnight carrier on 02 July 2014. Customers are asked to complete an inventory of goods to be returned which will ultimately be destroyed by the recalling company.

For general questions call the Carestream Customer Care Center at 800-933-8031 or 585-627-6557.

Item Information:

NSN (FSC-NIIN): 6525-014574992  
Part Number: 834 8658  
Nomenclature: KODAK DENTAL FILM; ULTRA-SPEED  
UI: EA  
Manufacturer: CARESTREAM HEALTH, INC.

LOT / SN:  
33301201  
33301202 33301203 34301501  
SEE\_NOTE\_1

**NOTES:**

NOTE 1:  
Lot #333012 part unknown  
NSN (FSC-NIIN): 6525-NS1  
Part Number: 811 0785  
Nomenclature: KODAK DENTAL FILM; INSIGHT  
UI: EA  
Manufacturer: CARESTREAM HEALTH, INC.

LOT / SN:  
53801303  
53801303-1 53901201 54601201  
54601201-1

NOTES:

POC Contact Information:

USAMMC-E

Email: [usarmy.rheinland-pfalz.medcom-usammce.list.dtd-quality-assurance@mail.mil](mailto:usarmy.rheinland-pfalz.medcom-usammce.list.dtd-quality-assurance@mail.mil)

Phone (COMM): 011-49-6331-86-7118/7181

Phone (DSN): 495-7118

Fax: 6218

US ARMY

Email: [usarmy.detrick.medcom-usamma.mbx.doc@mail.mil](mailto:usarmy.detrick.medcom-usamma.mbx.doc@mail.mil)

Phone (COMM): 301-619-4300

Phone (DSN): 343-4300

Fax: 301-619-4468

US AIR FORCE

Email: [afmoa.sgalc@us.af.mil](mailto:afmoa.sgalc@us.af.mil)

Phone (COMM): 301-619-4170

Phone (DSN): 343-4170

Fax: 301-619-2557

US NAVY

Email: [MMQC2@MED.NAVY.MIL](mailto:MMQC2@MED.NAVY.MIL)

Phone (COMM): 301-619-3085

Phone (DSN): 343-3085

Fax: 301-619-2473

Additional Message Recipients:

Please ensure widest dissemination of information to all interested entities.

Message Dissemination Authorization:

US AIR FORCE

AF Activities Will Take Action As Prescribed In Afi 41-209, Medical Logistics Support, Chapters 3 And 9. For MAJCOMS & NGB--This Msg Has Been Transmitted To All Designated Subordinate Medical Activities.

US ARMY

See Army Regulation (Ar) 40-61, 28 January 2005, Chapter 4, And The Department Of The Army Supply Bulletin (Sb 8-75-11) For Applicable Policies And Procedures.